



TECHNICAL DATA SHEET

GOWN FOR PATIENT - SHORT SLEEVE

(EU) 2017/745 Annex II-III Production Quality Assurance

GOWN FOR PATIENT - SHORT SLEEVE

► Product Description	► Protective equipment used by doctors or nurses for surgical operations to prevent the possible risk of infection.
► Product Class	► (EU) 2017/745 Medical Device Regulation – Class Business Rule I
► Manufacturer's Location	► Tio Medikal 2/20 st. No.:53 (Begos 3. North Entrance, 35400 Buca OSB/Buca/İzmir
► Purpose of usage	► These products are used as a protective barrier before, during and after surgical operations to prevent the risk of infection between patients, doctors, nurses or medical professionals in hospitals, clinics or intensive care units.

Quality

- Produced under ISO 13485:2016 and 13795-1 quality management standards.
- It has *CE certificate.

Bio-Compatibility

- Does not contain latex.

Related Standart

- ISO 13485:2016 Medical Devices Quality Management System
- 9001:2015 Quality Management System

- TS EN 13795 Surgical Clothing and Drapes

Shelf Life

- 5 years

Country of Origin

This disposable Patient Gown is completely made in Turkey



Shape 1. GOWN FOR PATIENT -
SHORT SLEEVE

1. GOWN FOR PATIENT - SHORTSLEEVE




Figure 1. GOWN FOR PATIENT - SHORT SLEEVE

Available Materials: SMS (Spunbond PP/Meltblown/Spunbond PP) 35-40 gr/m2

		DIMENSIONS (cm)					Pieces in Box	
REF . CODES		A	B	C	D	E	50×80×50	40×60×40
M	362.06.P01.02	115	105	29	30	140	100	50
L	363.06.P01.02	125	115	29	30	145	90	40
XL	364.06.P01.02	130	120	32	32	150	80	35
XXL	365.06.P01.02	135	125	32	34	160		

Tolerances: -+3% (ALL SIZES ARE PRODUCED & CUSTOMISATION)

		EN 13795 requirements			
Characteristic	Test Method	High performance		Results	Main fabric
		Critical product area	Less critical product area		
Resistance to microbial penetration - Dry (CFU)	EN ISO 22612	Not required	<300	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Resistance to microbial penetration - Wet (/B)	EN ISO 22610	6,0	Not required	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Cleanliness - Microbial (CFU/100cm²)	EN ISO 11737-1	<300	<300	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Cleanliness - Particulate matter (IPM)	EN ISO 9073-10	<3,5	<3,5	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Linting (Log10 (lint count))	EN ISO 9073-10	<4,0	<4,0	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Resistance to liquid penetration	EN 13795	≥ 100	≥ 10	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Bursting strength - dry	EN 13795	≥ 40	≥ 40	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Bursting strength - wet	EN 13795	Not required	≥ 40	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Tensile strength - dry	EN 13795	≥ 20	≥ 20	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Tensile strength - wet	EN 13795	Not required	≥ 20	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven

BIOCOMPATIBILITY

Biocompatibility studies has been done, and test results are given in the below table. Test results are within limits.

Table 31 Biocompatibility Study Documentation

Test Name		Test Method	Test facility	Report No	Report Date	Result
Cytotoxicity		ISO10993-5:2010	HACETTEPE UNIVERSITY	ARGEDS-2016/37	01.08.2016	Confirmed
Sensitizasyon Sensitization		ISO10993-10:2014	HACETTEPE UNIVERSITY	ARGEDS-2016/37	24.10.2016	Confirmed
Cilt İritasyon Skin Irritation		ISO10993-10:2014	HACETTEPE UNIVERSITY	ARGEDS-2016/37	11.06.2016	Confirmed

Instruction for use:

1. The package of product shall be opened in sterile and aseptic conditions.
2. For a clean peel, open the package from the direction of the arrow slowly .
3. There are labels or marks on the drapes, which helps the user to follow the patient direction and unfolding directions.
4. For the incision film or adhesive tape on the drapes,first of all peel the paper carrier and fix the drape on the operational area of the patient. After then unfold the drape by following the directions.
5. Surgical drapes are ready for the operation, when they are completely unfolded.

	Do not use if package is damaged		Manufacturer		Single Use (Do not re-use)		Use by date
	Do not expose the product to sunlight.		Catalogue number		Consult instructions for use		Temperature Limitation
	Keep Dry		Manufacturing Date		Caution		CE Marking
	Unique device identifier		Batch Code		Medical device		

